Cleaning and Sanitization Record for Incubator

Equipment Name:

Equipment No:

Range:

Cleaning Frequency:

DATE	TIME	DISINFECTANT USED	CONCENTRATION	CLEANED BY	CHECKED BY	NOTE(S)

Buy CAPTiCLEAN™ on captivatebio.com | Cat. No. CBL003

Call (617) 607-4017 | **Email** orders@captivatebio.com | **Shop** www.captivatebio.com

©2025-2026 Captivate Bio, LLC. All trademarks are the property of Captivate Bio and/or their respective owners. All rights reserved. For research use only. Version: 1025-1

CAPTIVATE BIO

Maintenance Checklist

The below maintenance schedule is provided as a guide and may be adjusted per lab protocols.

DAILY

- ☐ Check for spills, clean with 70% isopropanol
- ☐ Add water to tray as needed

WEEKLY

- ☐ Empty water tray & clean it
- ☐ Record temp, humidity, and CO_2
- ☐ Refill water tray
- ☐ Add 10mL of CAPTiCLEAN per liter of water

MONTHLY

- ☐ Fully empty incubator
- ☐ Wipe interior with 70% isopropanol
- ☐ Clean exterior (door handles etc)
- ☐ Remove any dust from top and back
- ☐ Inspect and clean door gaskets
- ☐ Fully clean and refill water trays
- ☐ Add 10mL of CAPTiCLEAN per liter of water

ANNUALLY

- ☐ Calibrate temp, humidity, and CO_2 sensors
- ☐ Replace HEPA filters
- ☐ Schedule routine maintenance visits
- ☐ Stock up on CAPTiCLEAN